
Midmark M11 - v1.0.0
Total Cycles: 3725

Name: RITTER 1
Sterilizer ID: V2178198
09/06/2023 11:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.2 F 0.0
2:00 160.3 F 1.6
4:00 175.9 F 4.4
6:00 196.8 F 8.3
8:00 214.3 F 13.2
10:00 228.0 F 7.4
12:00 233.9 F 12.2
14:00 236.9 F 9.3
16:00 239.0 F 9.9
18:00 247.1 F 13.7
20:00 258.5 F 19.9
22:00 268.9 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.7 F 29.2
2:00 271.6 F 28.5
3:00 271.8 F 29.0
4:00 272.9 F 29.1

Min 271.1 F 28.4
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 220.1 F 0.0
10:00 192.9 F 0.0
15:00 190.1 F 0.0
20:00 186.4 F 0.0
25:00 183.6 F 0.0
30:00 181.4 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:01:59
CYCLE COMPLETE

09/06/2023 12:13 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____