
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000720
7:10 22/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:11
PRESSURIZING 2:20
STERILIZING 3:12
136.0 C 226kPa 3:12
136.4 C 232kPa 3:42
136.4 C 232kPa 4:12
136.1 C 230kPa 4:42
136.1 C 229kPa 5:12
136.3 C 230kPa 5:42
136.4 C 232kPa 6:12
136.3 C 231kPa 6:42
136.3 C 231kPa 6:42
VENTING 6:42
AIR DRYING 7:29
CYCLE COMPLETE 11:40

Digital Signature #
AAF934E2B4B119B9
