
Midmark M11 - v1.0.0
Total Cycles: 4952

Name: RITTER 1
Sterilizer ID: V2178198
10/02/2024 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 163.0 F 0.0
2:00 162.4 F 2.3
4:00 174.8 F 5.6
6:00 193.9 F 9.2
8:00 213.5 F 13.0
10:00 221.5 F 6.2
12:00 231.1 F 10.5
14:00 233.5 F 7.8
16:00 240.1 F 11.7
18:00 237.6 F 9.4
20:00 249.5 F 15.2
22:00 260.0 F 21.0
24:00 269.3 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.5 F 29.2
2:00 271.9 F 28.7
3:00 271.5 F 28.7
4:00 272.4 F 29.2

Min 271.3 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 206.3 F 0.0
10:00 209.2 F 0.0
15:00 207.6 F 0.0
20:00 208.4 F 0.0
25:00 209.1 F 0.0
30:00 213.3 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:04:03
CYCLE COMPLETE

10/02/2024 10:48 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____