
Midmark M11 - v1.0.0
Total Cycles: 4961

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2024 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 132.7 F 0.0
2:00 138.3 F 0.7
4:00 161.7 F 1.9
6:00 197.5 F 5.8
8:00 217.0 F 9.9
10:00 232.2 F 8.8
12:00 229.2 F 8.3
14:00 241.1 F 12.8
16:00 237.5 F 9.5
18:00 239.1 F 9.6
20:00 246.3 F 13.2
22:00 257.5 F 19.2
24:00 267.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 271.5 F 28.6
2:00 272.8 F 29.2
3:00 271.9 F 28.6
4:00 272.0 F 28.9

Min 270.7 F 27.9
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.5 F 0.0
5:00 228.3 F 0.0
10:00 184.4 F 0.0
15:00 180.1 F 0.0
20:00 178.1 F 0.0
25:00 178.2 F 0.0
30:00 178.1 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 25:18
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:03:56
CYCLE COMPLETE

10/03/2024 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____