
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002251
12:23 13/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:41
PRESSURIZING 2:45
STERILIZING 3:47
136.0 C 222kPa 3:47
136.3 C 224kPa 4:17
136.4 C 225kPa 4:47
136.4 C 222kPa 5:17
136.4 C 225kPa 5:47
136.4 C 225kPa 6:17
136.2 C 222kPa 6:47
136.4 C 225kPa 7:17
136.4 C 225kPa 7:17
VENTING 7:17
AIR DRYING 8:12
CYCLE COMPLETE 16:51

Digital Signature #
BDB6D91B4DCE58CA
