

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 002581  
7:58 21/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:50  
PRESSURIZING 1:57  
STERILIZING 2:57  
136.1 C 220kPa 2:57  
136.3 C 223kPa 3:27  
136.4 C 224kPa 3:57  
136.4 C 222kPa 4:27  
136.2 C 221kPa 4:57  
136.3 C 223kPa 5:27  
136.3 C 224kPa 5:57  
136.1 C 222kPa 6:27  
136.1 C 222kPa 6:27  
VENTING 6:27  
AIR DRYING 7:17  
CYCLE COMPLETE 19:25  
-----

Digital Signature #  
25472555920C6A93  
-----