
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000711
10:30 20/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:14
STERILIZING 4:19
136.0 C 227kPa 4:19
136.3 C 232kPa 4:49
136.2 C 231kPa 5:19
136.4 C 233kPa 5:49
136.0 C 232kPa 6:19
136.2 C 231kPa 6:49
136.4 C 233kPa 7:19
136.3 C 232kPa 7:49
136.3 C 232kPa 7:49
VENTING 7:49
AIR DRYING 8:45
CYCLE COMPLETE 8:48

Digital Signature #
7899476E3165D5A4
