
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000322
14:06 04/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:20
STERILIZING 4:37
136.0 C 224kPa 4:37
136.3 C 232kPa 5:07
136.3 C 232kPa 5:37
136.2 C 231kPa 6:07
136.4 C 231kPa 6:37
136.4 C 233kPa 7:07
136.3 C 232kPa 7:37
136.4 C 232kPa 8:07
136.4 C 232kPa 8:07
VENTING 8:07
AIR DRYING 9:01
CYCLE COMPLETE 24:01

Digital Signature #
59ACE76F3E5184A0
