
Midmark M11 - v1.0.5
Total Cycles: 1479

Name: RITTER2
Sterilizer ID: V2313228
09/12/2022 04:22 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.0 F 0.0
2:00 189.6 F 4.0
4:00 203.9 F 9.8
6:00 219.6 F 5.7
8:00 235.2 F 11.8
10:00 237.4 F 10.1
12:00 236.8 F 9.3
14:00 252.6 F 16.9
16:00 265.4 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.1 F 28.8
2:00 272.7 F 29.3
3:00 272.0 F 29.3
4:00 272.0 F 28.8

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.6 F 0.0
5:00 219.3 F 0.0
10:00 236.7 F 0.0
15:00 230.8 F 0.0
20:00 220.3 F 0.0
25:00 215.0 F 0.0
30:00 212.7 F 0.0
Drying Complete

FILLING: 2:13
HEATING: 17:34
STERILIZING: 4:00
VENTING: 2:55
DRYING: 30:00
TOTAL CYCLE: 00:56:42
CYCLE COMPLETE

09/12/2022 05:18 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____