

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000761  
7:04 28/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:57  
PRESSURIZING 2:07  
STERILIZING 2:58  
136.0 C 220kPa 2:58  
136.0 C 230kPa 3:28  
136.4 C 233kPa 3:58  
136.3 C 231kPa 4:28  
136.4 C 233kPa 4:58  
136.4 C 232kPa 5:28  
136.4 C 233kPa 5:58  
136.3 C 231kPa 6:28  
136.3 C 231kPa 6:28  
VENTING 6:28  
AIR DRYING 7:15  
CYCLE COMPLETE 10:09  
-----

Digital Signature #  
5B5CD57C9BA9A322  
-----