
Midmark M11 - v1.0.5
Total Cycles: 4366

Name: RITTER2
Sterilizer ID: V2313228
02/20/2025 05:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.3 F 0.0
2:00 157.3 F 2.0
4:00 170.8 F 5.5
6:00 201.8 F 9.5
8:00 224.9 F 8.1
10:00 222.6 F 7.5
12:00 235.1 F 12.1
14:00 235.3 F 9.3
16:00 240.0 F 10.4
18:00 246.5 F 13.9
20:00 257.7 F 20.0
22:00 267.9 F 26.5
24:00 273.1 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.8
1:00 272.1 F 29.7
2:00 272.2 F 29.8
3:00 272.2 F 29.9
4:00 272.3 F 30.1

Min 271.9 F 29.7
Max 272.5 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 205.2 F 0.0
10:00 207.5 F 0.0
15:00 202.8 F 0.0
20:00 193.6 F 0.0
25:00 194.6 F 0.0
30:00 196.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 24:13
STERILIZING: 4:00
VENTING: 4:01
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

02/20/2025 06:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____