
Midmark M11 - v1.0.0
Total Cycles: 713

Name: MIDMARK
Sterilizer ID: V2178198
03/01/2021 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.1 F 0.0
2:00 176.0 F 3.7
4:00 186.3 F 8.4
6:00 211.5 F 8.4
8:00 227.9 F 9.4
10:00 231.7 F 8.3
12:00 243.9 F 13.0
14:00 242.4 F 11.7
16:00 256.0 F 18.9
18:00 268.2 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.5 F 28.9
2:00 272.6 F 29.3
3:00 272.0 F 28.6
4:00 273.1 F 29.4

Min 271.5 F 28.4
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 233.4 F 0.0
10:00 229.4 F 0.0
15:00 214.0 F 0.0
Dry Cycle Aborted

FILLING: 1:15
HEATING: 19:09
STERILIZING: 4:00
VENTING: 2:54
DRYING: 15:27
TOTAL CYCLE: 00:42:46

03/01/2021 12:34 PM
ACCEPTED: X
REJECTED:
Operator: U01