
Midmark M11 - v1.0.0
Total Cycles: 1217

Name: RITTER 1
Sterilizer ID: V2178198
09/29/2021 09:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.3 F 0.0
2:00 181.8 F 3.2
4:00 193.3 F 8.1
6:00 218.2 F 10.2
8:00 226.5 F 9.8
10:00 233.5 F 7.9
12:00 243.0 F 13.1
14:00 242.4 F 11.6
16:00 255.6 F 18.4
18:00 267.0 F 25.4
20:00 272.7 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.7 F 29.1
2:00 272.6 F 29.5
3:00 271.8 F 28.7
4:00 272.9 F 29.4

Min 271.6 F 28.7
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.8 F 0.0
5:00 246.6 F 0.0
10:00 237.3 F 0.0
15:00 228.9 F 0.0
Dry Cycle Aborted

FILLING: 1:36
HEATING: 20:11
STERILIZING: 4:00
VENTING: 3:04
DRYING: 19:40
TOTAL CYCLE: 00:48:31

09/29/2021 10:00 AM
ACCEPTED: X
REJECTED:
Operator: U01