
Midmark M11 - v1.0.0
Total Cycles: 1864

Name: RITTER 1
Sterilizer ID: V2178198
04/25/2022 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.2 F 0.0
2:00 164.5 F 2.6
4:00 181.5 F 6.5
6:00 201.6 F 11.6
8:00 216.3 F 5.3
10:00 230.2 F 10.4
12:00 233.5 F 7.4
14:00 243.7 F 13.2
16:00 242.9 F 11.7
18:00 256.0 F 18.5
20:00 267.5 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.5 F 29.0
2:00 271.4 F 28.4
3:00 272.5 F 29.1
4:00 271.6 F 28.3

Min 271.3 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.2 F 0.0
5:00 200.1 F 0.0
10:00 210.6 F 0.0
15:00 209.1 F 0.0
20:00 211.4 F 0.0
25:00 224.3 F 0.0
30:00 230.7 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:20
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:59:45
CYCLE COMPLETE

04/25/2022 03:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____