
Midmark M11 - v1.0.0
Total Cycles: 2987

Name: RITTER 1
Sterilizer ID: V2178198
03/09/2023 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 185.3 F 0.0
2:00 180.0 F 3.2
4:00 189.3 F 7.7
6:00 209.5 F 13.0
8:00 223.3 F 7.6
10:00 236.5 F 12.9
12:00 237.6 F 10.1
14:00 235.9 F 8.1
16:00 250.2 F 15.3
18:00 262.3 F 22.2
20:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.5 F 29.1
2:00 271.9 F 28.4
3:00 271.6 F 28.5
4:00 272.6 F 29.1

Min 271.4 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.3 F 0.0
5:00 224.0 F 0.0
10:00 234.6 F 0.0
15:00 228.1 F 0.0
20:00 224.2 F 0.0
25:00 223.4 F 0.0
30:00 234.5 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 20:19
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:47
CYCLE COMPLETE

03/09/2023 11:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____