
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000526
10:58 04/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:41
PRESSURIZING 3:48
STERILIZING 4:54
136.0 C 227kPa 4:54
136.4 C 230kPa 5:25
136.3 C 232kPa 5:55
136.4 C 229kPa 6:25
136.2 C 231kPa 6:55
136.3 C 231kPa 7:25
136.3 C 230kPa 7:55
136.4 C 233kPa 8:25
136.4 C 233kPa 8:25
VENTING 8:25
AIR DRYING 9:19
CYCLE COMPLETE 24:19

Digital Signature #
EE1D200EB488C4EF
