
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
1.9uS / 1.2ppm
CYCLE NUMBER 000037
16:22 10/08/2020
HEAVY DUTY UNWRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 1:35
PRESSURIZING 3:35
STERILIZING 4:30
136.0 C 222kPa 4:30
136.4 C 227kPa 5:00
136.4 C 227kPa 5:30
136.2 C 225kPa 6:00
136.1 C 224kPa 6:30
136.4 C 226kPa 7:00
136.4 C 225kPa 7:30
136.5 C 227kPa 8:01
135.8 C 222kPa 8:31
136.4 C 225kPa 9:01
136.4 C 227kPa 9:31
136.4 C 227kPa 10:01
136.4 C 226kPa 10:30
VENTING 10:30
AIR DRYING 11:21
CYCLE COMPLETE 13:21

Digital Signature #
C53786E9CD2EDCA3
