

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3794

Name: RITTER 1  
Sterilizer ID: V2178198  
09/19/2023 11:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
2:00 188.0 F 1.4  
4:00 189.7 F 5.1  
6:00 214.4 F 9.9  
8:00 228.1 F 7.3  
10:00 230.8 F 9.6  
12:00 236.9 F 9.5  
14:00 243.0 F 11.9  
16:00 240.1 F 10.3  
18:00 252.9 F 16.6  
20:00 264.0 F 23.1  
22:00 271.6 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.3 F 28.5  
2:00 272.1 F 28.7  
3:00 272.0 F 28.2  
4:00 272.4 F 28.9

Min 271.4 F 28.0  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.3 F 0.0  
5:00 223.8 F 0.0  
10:00 229.7 F 0.0  
15:00 230.8 F 0.0  
20:00 223.6 F 0.0  
25:00 215.1 F 0.0  
30:00 212.5 F 0.0  
Drying Complete

FILLING: 1:48  
HEATING: 22:03  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 01:01:20  
CYCLE COMPLETE

09/19/2023 12:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_