
Midmark M11 - v1.0.5
Total Cycles: 27

Name: RITTER2
Sterilizer ID: V2313228
04/27/2021 02:32 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.7 F 0.0
2:00 132.4 F 1.8
4:00 151.0 F 5.1
6:00 179.9 F 9.5
8:00 205.5 F 6.5
10:00 222.1 F 8.2
12:00 233.8 F 13.6
14:00 239.1 F 11.3
16:00 238.1 F 10.0
18:00 251.4 F 16.5
20:00 262.9 F 23.2
22:00 272.8 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.3
1:00 272.9 F 29.8
2:00 271.5 F 29.6
3:00 272.2 F 29.7
4:00 271.5 F 29.5

Min 271.4 F 29.2
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 176.3 F 0.0
5:00 223.4 F 0.0
10:00 192.8 F 0.0
15:00 201.6 F 0.0
20:00 192.2 F 0.0
25:00 191.4 F 0.0
30:00 198.8 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 22:19
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:00:47
CYCLE COMPLETE

04/27/2021 03:32 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____