
Midmark M11 - v1.0.0
Total Cycles: 1536

Name: RITTER 1
Sterilizer ID: V2178198
01/19/2022 10:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.1 F 0.0
2:00 102.3 F 0.0
4:00 125.3 F 1.3
6:00 153.5 F 4.3
8:00 177.9 F 8.0
10:00 196.2 F 12.3
12:00 210.8 F 4.5
14:00 222.8 F 8.8
16:00 233.8 F 13.1
18:00 234.5 F 9.3
20:00 239.5 F 10.2
22:00 244.4 F 12.7
24:00 255.9 F 18.7
26:00 266.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.6
1:00 272.6 F 29.3
2:00 271.7 F 28.6
3:00 272.1 F 29.2
4:00 272.4 F 29.1

Min 271.3 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 218.3 F 0.0
10:00 231.7 F 0.0
15:00 225.6 F 0.0
20:00 226.8 F 0.0
25:00 233.2 F 0.0
30:00 229.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 27:33
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:05:41
CYCLE COMPLETE

01/19/2022 11:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____