

-----  
Midmark M11 - v1.0.5  
Total Cycles: 794

Name: RITTER2  
Sterilizer ID: V2313228  
02/14/2022 11:55 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.2 F 0.0  
2:00 164.8 F 3.1  
4:00 189.8 F 7.9  
6:00 211.2 F 13.5  
8:00 225.3 F 8.8  
10:00 234.5 F 9.8  
12:00 242.7 F 13.1  
14:00 243.6 F 12.5  
16:00 257.1 F 19.7  
18:00 269.2 F 27.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.0  
1:00 272.5 F 29.4  
2:00 271.8 F 29.4  
3:00 272.4 F 29.3  
4:00 271.7 F 29.4  
  
Min 271.4 F 28.9  
Max 272.8 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 186.7 F 0.0  
5:00 216.3 F 0.0  
10:00 215.2 F 0.0  
15:00 212.8 F 0.0  
20:00 218.3 F 0.0  
25:00 230.0 F 0.0  
30:00 220.3 F 0.0  
Drying Complete

FILLING: 1:39  
HEATING: 18:60  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:47  
CYCLE COMPLETE

02/14/2022 12:52 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_