
Midmark M11 - v1.0.5
Total Cycles: 2368

Name: RITTER2
Sterilizer ID: V2313228
05/15/2023 02:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.8 F 0.0
2:00 171.4 F 3.2
4:00 182.7 F 7.4
6:00 203.7 F 12.6
8:00 222.0 F 7.3
10:00 234.1 F 12.6
12:00 235.9 F 10.2
14:00 235.5 F 8.5
16:00 249.7 F 15.8
18:00 261.8 F 22.6
20:00 272.7 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.1
1:00 272.7 F 29.9
2:00 271.8 F 29.1
3:00 272.2 F 29.9
4:00 272.4 F 29.6

Min 271.4 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.7 F 0.0
5:00 212.8 F 0.0
10:00 220.9 F 0.0
15:00 220.0 F 0.0
20:00 218.2 F 0.0
25:00 219.4 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:24
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

05/15/2023 03:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____