
Midmark M11 - v1.0.5
Total Cycles: 2272

Name: RITTER2
Sterilizer ID: V2313228
04/20/2023 07:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.8 F 0.0
2:00 79.3 F 0.0
4:00 87.3 F 0.0
6:00 119.1 F 0.1
8:00 154.5 F 3.0
10:00 177.0 F 6.3
12:00 195.8 F 9.9
14:00 210.0 F 9.2
16:00 220.9 F 8.0
18:00 232.0 F 12.7
20:00 233.8 F 9.4
22:00 238.1 F 10.1
24:00 244.9 F 13.2
26:00 256.4 F 19.3
28:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.0 F 29.5
2:00 272.2 F 29.2
3:00 271.3 F 29.0
4:00 272.5 F 29.7

Min 271.1 F 28.7
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.7 F 0.0
5:00 204.1 F 0.0
10:00 215.6 F 0.0
15:00 221.1 F 0.0
20:00 217.7 F 0.0
25:00 217.7 F 0.0
30:00 216.1 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 29:25
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:07:19
CYCLE COMPLETE

04/20/2023 08:18 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____