

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1363

Name: RITTER 1  
Sterilizer ID: V2178198  
11/11/2021 01:53 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.5 F 0.0  
2:00 193.6 F 3.7  
4:00 204.9 F 8.7  
6:00 230.8 F 7.8  
8:00 236.9 F 10.1  
10:00 233.2 F 7.8  
12:00 245.1 F 12.5  
14:00 245.2 F 12.6  
16:00 258.5 F 19.6  
18:00 270.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 272.2 F 28.3  
1:00 272.6 F 28.6  
2:00 272.6 F 28.8  
3:00 272.0 F 28.9  
4:00 271.6 F 28.2  
  
Min 271.4 F 28.2  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.4 F 0.0  
5:00 222.7 F 0.0  
10:00 201.6 F 0.0  
15:00 188.5 F 0.0  
20:00 218.0 F 0.0  
25:00 197.8 F 0.0  
30:00 194.9 F 0.0  
Drying Complete

FILLING: 2:27  
HEATING: 19:24  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:60  
CYCLE COMPLETE

11/11/2021 02:56 PM  
ACCEPTED: X  
REJECTED:  
Operator: U04