
Midmark M11 - v1.0.0
Total Cycles: 4549

Name: RITTER 1
Sterilizer ID: V2178198
05/14/2024 08:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 77.2 F 0.0
4:00 98.9 F 0.2
6:00 128.9 F 2.0
8:00 155.8 F 4.9
10:00 182.0 F 8.4
12:00 207.7 F 12.2
14:00 211.5 F 4.0
16:00 223.6 F 7.3
18:00 234.0 F 11.0
20:00 229.6 F 6.8
22:00 240.3 F 11.1
24:00 235.0 F 8.2
26:00 248.8 F 14.5
28:00 259.9 F 20.7
30:00 270.0 F 27.3

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.3 F 28.5
2:00 272.5 F 29.0
3:00 271.7 F 28.3
4:00 272.4 F 28.7

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 204.0 F 0.0
10:00 211.7 F 0.0
15:00 209.2 F 0.0
20:00 223.8 F 0.0
25:00 229.8 F 0.0
30:00 226.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:53
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:09:07
CYCLE COMPLETE

05/14/2024 09:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____