
Midmark M11 - v1.0.0
Total Cycles: 4770

Name: RITTER 1
Sterilizer ID: V2178198
07/26/2024 12:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.8 F 0.0
2:00 171.3 F 3.4
4:00 181.0 F 7.1
6:00 202.4 F 11.5
8:00 216.1 F 5.0
10:00 227.7 F 9.5
12:00 235.7 F 9.7
14:00 239.9 F 11.3
16:00 237.8 F 9.4
18:00 250.8 F 15.7
20:00 261.9 F 22.0
22:00 272.1 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.0
2:00 271.7 F 28.4
3:00 271.6 F 28.7
4:00 272.5 F 29.0

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.8 F 0.0
5:00 215.5 F 0.0
10:00 240.8 F 0.0
15:00 236.4 F 0.0
20:00 231.7 F 0.0
25:00 224.4 F 0.0
30:00 223.4 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

07/26/2024 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____