

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1534

Name: RITTER2  
Sterilizer ID: V2313228  
09/26/2022 12:13 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 133.2 F 0.0  
2:00 134.8 F 0.5  
4:00 153.7 F 3.5  
6:00 177.1 F 7.5  
8:00 202.0 F 12.6  
10:00 218.1 F 7.0  
12:00 231.2 F 11.9  
14:00 232.4 F 9.0  
16:00 238.6 F 10.4  
18:00 245.4 F 13.9  
20:00 257.7 F 20.6  
22:00 269.0 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 29.6  
1:00 272.7 F 30.1  
2:00 271.3 F 29.4  
3:00 272.3 F 30.3  
4:00 271.8 F 29.7

Min 271.2 F 29.4  
Max 272.8 F 30.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.2 F 0.0  
5:00 215.6 F 0.0  
10:00 228.1 F 0.0  
15:00 229.0 F 0.0  
20:00 228.6 F 0.0  
25:00 229.3 F 0.0  
30:00 226.8 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 23:06  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:17  
CYCLE COMPLETE

09/26/2022 01:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_