
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008346
9:54 25/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:04
STERILIZING 4:03
136.0 C 220kPa 4:03
136.4 C 223kPa 4:33
136.2 C 221kPa 5:03
136.4 C 224kPa 5:33
136.3 C 223kPa 6:03
136.4 C 222kPa 6:33
136.4 C 224kPa 7:03
136.3 C 218kPa 7:33
136.3 C 218kPa 7:33
VENTING 7:33
AIR DRYING 8:34
CYCLE COMPLETE 19:18

Digital Signature #
912E23ACCAEC9B22
