
Midmark M11 - v1.0.0
Total Cycles: 2165

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 07:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.9 F 0.0
2:00 75.7 F 0.0
4:00 82.3 F 0.0
6:00 108.3 F 0.0
8:00 140.8 F 0.1
10:00 165.7 F 3.0
12:00 188.8 F 6.3
14:00 214.9 F 9.4
16:00 230.3 F 12.7
18:00 222.8 F 5.8
20:00 232.8 F 9.8
22:00 240.8 F 10.8
24:00 239.3 F 10.7
26:00 236.4 F 7.9
28:00 248.2 F 14.4
30:00 259.0 F 20.3
32:00 268.8 F 26.6
34:00 271.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.7 F 29.2
2:00 271.8 F 28.9
3:00 272.5 F 29.0
4:00 271.9 F 29.0

Min 271.7 F 28.6
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 201.5 F 0.0
10:00 195.1 F 0.0
15:00 193.2 F 0.0
20:00 224.7 F 0.0
25:00 230.3 F 0.0
30:00 233.8 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 34:01
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:11:57
CYCLE COMPLETE

07/25/2022 08:42 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____