
Midmark M11 - v1.0.0
Total Cycles: 4972

Name: RITTER 1
Sterilizer ID: V2178198
10/07/2024 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 101.5 F 0.0
2:00 103.8 F 0.2
4:00 130.1 F 1.0
6:00 159.3 F 2.1
8:00 180.3 F 4.9
10:00 198.0 F 8.0
12:00 208.1 F 11.6
14:00 221.0 F 4.3
16:00 225.6 F 8.5
18:00 235.1 F 12.7
20:00 234.7 F 9.2
22:00 240.5 F 10.8
24:00 244.8 F 12.9
26:00 256.2 F 18.7
28:00 265.9 F 24.6
30:00 271.1 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.4
1:00 271.4 F 28.6
2:00 271.9 F 29.0
3:00 272.2 F 29.2
4:00 272.6 F 29.6

Min 271.0 F 28.3
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.4 F 0.0
5:00 202.6 F 0.0
10:00 221.7 F 0.0
15:00 226.3 F 0.0
20:00 233.9 F 0.0
25:00 228.5 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 30:04
STERILIZING: 4:00
VENTING: 4:06
DRYING: 30:00
TOTAL CYCLE: 01:08:58
CYCLE COMPLETE

10/07/2024 02:58 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____