

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3030

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2023 08:13 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 80.3 F 0.0  
2:00 103.8 F 0.0  
4:00 110.8 F 0.0  
6:00 145.9 F 0.4  
8:00 169.9 F 3.7  
10:00 186.0 F 7.5  
12:00 202.4 F 11.4  
14:00 215.0 F 4.3  
16:00 225.3 F 8.1  
18:00 235.9 F 12.4  
20:00 234.6 F 8.7  
22:00 241.3 F 10.9  
24:00 245.0 F 12.7  
26:00 256.9 F 19.0  
28:00 267.8 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.3  
1:00 272.2 F 28.5  
2:00 272.4 F 28.7  
3:00 272.6 F 28.9  
4:00 272.2 F 28.9

Min 271.5 F 28.1  
Max 272.7 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 173.4 F 0.0  
5:00 223.0 F 0.0  
10:00 227.1 F 0.0  
15:00 228.8 F 0.0  
20:00 226.9 F 0.0  
25:00 231.6 F 0.0  
30:00 230.3 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 29:08  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:06:59  
CYCLE COMPLETE

03/24/2023 09:19 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_