
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000623
14:29 01/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:17
STERILIZING 4:39
136.4 C 231kPa 4:39
136.4 C 233kPa 5:09
136.4 C 233kPa 5:39
136.4 C 232kPa 6:09
136.4 C 232kPa 6:39
136.4 C 233kPa 7:09
136.4 C 233kPa 7:39
136.4 C 232kPa 8:09
136.4 C 232kPa 8:09
VENTING 8:09
AIR DRYING 9:04
CYCLE COMPLETE 24:04

Digital Signature #
7E1D6A81C33AB797
