
Midmark M11 - v1.0.0
Total Cycles: 1914

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2022 11:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.6 F 0.0
2:00 178.4 F 3.3
4:00 184.6 F 7.5
6:00 204.4 F 12.7
8:00 225.5 F 7.1
10:00 237.0 F 12.3
12:00 235.3 F 9.2
14:00 240.0 F 9.2
16:00 247.6 F 14.1
18:00 259.9 F 20.8
20:00 271.0 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.0
2:00 271.9 F 28.3
3:00 271.8 F 28.7
4:00 272.6 F 28.9

Min 271.3 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.6 F 0.0
5:00 228.5 F 0.0
10:00 225.0 F 0.0
15:00 230.6 F 0.0
20:00 235.3 F 0.0
25:00 227.0 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:46
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:09
CYCLE COMPLETE

05/05/2022 12:36 PM
ACCEPTED: X
REJECTED:
Operator: U02