
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000414
14:24 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:05
STERILIZING 4:02
136.4 C 225kPa 4:02
136.2 C 222kPa 4:32
136.2 C 223kPa 5:02
136.3 C 223kPa 5:32
136.4 C 224kPa 6:02
136.4 C 222kPa 6:32
136.1 C 222kPa 7:02
136.3 C 225kPa 7:32
136.3 C 225kPa 7:32
VENTING 7:32
AIR DRYING 8:24
CYCLE COMPLETE 10:38

Digital Signature #
EE4EEA468AE40AC4
