
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001084
16:07 18/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:17
PRESSURIZING 3:22
STERILIZING 4:19
136.0 C 220kPa 4:19
136.2 C 222kPa 4:49
136.4 C 226kPa 5:19
136.3 C 225kPa 5:49
136.1 C 223kPa 6:19
136.4 C 225kPa 6:49
136.4 C 223kPa 7:19
136.3 C 223kPa 7:49
136.3 C 223kPa 7:49
VENTING 7:49
AIR DRYING 8:41
CYCLE COMPLETE 23:41

Digital Signature #
6369E5B270168C2B
