
Midmark M11 - v1.0.5
Total Cycles: 4341

Name: RITTER2
Sterilizer ID: V2313228
02/07/2025 11:58 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 115.2 F 0.0
2:00 115.3 F 0.7
4:00 120.9 F 2.6
6:00 149.5 F 5.3
8:00 172.4 F 8.8
10:00 189.9 F 12.6
12:00 214.8 F 5.1
14:00 222.6 F 8.7
16:00 233.8 F 12.7
18:00 232.6 F 9.0
20:00 241.6 F 12.0
22:00 242.2 F 12.2
24:00 253.2 F 17.9
26:00 263.7 F 23.9
28:00 272.9 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 272.3 F 30.3
2:00 271.5 F 29.5
3:00 271.8 F 30.3
4:00 272.4 F 30.0

Min 271.2 F 29.5
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 226.9 F 0.0
10:00 205.6 F 0.0
15:00 205.2 F 0.0
20:00 189.1 F 0.0
25:00 183.3 F 0.0
30:00 193.2 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 28:24
STERILIZING: 4:00
VENTING: 3:55
DRYING: 30:00
TOTAL CYCLE: 01:07:13
CYCLE COMPLETE

02/07/2025 01:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____