
Midmark M11 - v1.0.5
Total Cycles: 3731

Name: RITTER2
Sterilizer ID: V2313228
06/24/2024 02:18 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 93.1 F 0.0
2:00 93.1 F 0.2
4:00 104.0 F 1.6
6:00 140.5 F 5.1
8:00 170.3 F 9.4
10:00 193.5 F 6.9
12:00 213.8 F 5.8
14:00 225.9 F 9.9
16:00 232.1 F 9.7
18:00 238.9 F 11.2
20:00 236.5 F 9.2
22:00 245.1 F 13.0
24:00 257.8 F 19.8
26:00 269.0 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.6 F 28.6
2:00 271.9 F 28.5
3:00 272.3 F 28.8
4:00 272.7 F 29.1

Min 271.6 F 28.5
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 162.6 F 0.0
5:00 204.4 F 0.0
10:00 207.6 F 0.0
15:00 206.5 F 0.0
20:00 210.1 F 0.0
25:00 211.3 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 26:59
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:04:51
CYCLE COMPLETE

06/24/2024 03:22 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____