

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2287

Name: RITTER 1  
Sterilizer ID: V2178198  
09/02/2022 04:32 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 188.9 F 0.0  
2:00 191.6 F 1.9  
4:00 215.7 F 7.6  
6:00 231.8 F 13.4  
8:00 233.2 F 9.5  
10:00 238.6 F 8.6  
12:00 245.5 F 13.5  
14:00 244.2 F 12.4  
16:00 257.6 F 19.2  
18:00 268.9 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.2  
1:00 272.4 F 28.7  
2:00 271.5 F 28.2  
3:00 272.3 F 28.8  
4:00 271.5 F 28.1  
  
Min 271.2 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.6 F 0.0  
5:00 208.3 F 0.0  
10:00 219.9 F 0.0  
15:00 227.0 F 0.0  
20:00 232.3 F 0.0  
25:00 225.9 F 0.0  
30:00 224.4 F 0.0  
Drying Complete

FILLING: 2:01  
HEATING: 19:06  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:58:20  
CYCLE COMPLETE

09/02/2022 05:30 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_