
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 006509
11:35 20/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:49
STERILIZING 2:40
136.0 C 220kPa 2:40
136.4 C 223kPa 3:10
136.4 C 223kPa 3:40
136.4 C 225kPa 4:10
136.4 C 224kPa 4:40
136.4 C 223kPa 5:10
136.4 C 223kPa 5:40
136.4 C 220kPa 6:10
136.4 C 220kPa 6:10
VENTING 6:10
AIR DRYING 7:02
CYCLE COMPLETE 37:02

Digital Signature #
F17A062A19002730
