

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3331

Name: RITTER 1  
Sterilizer ID: V2178198  
06/05/2023 07:27 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 73.6 F 0.0  
2:00 74.9 F 0.0  
4:00 78.9 F 0.0  
6:00 111.3 F 0.0  
8:00 141.3 F 1.3  
10:00 183.6 F 4.2  
12:00 189.3 F 7.5  
14:00 210.4 F 11.3  
16:00 220.5 F 4.9  
18:00 226.8 F 8.5  
20:00 237.6 F 12.7  
22:00 235.9 F 8.9  
24:00 241.1 F 10.8  
26:00 244.6 F 12.3  
28:00 256.1 F 18.3  
30:00 266.5 F 24.6  
32:00 272.9 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.5 F 28.6  
2:00 272.7 F 29.0  
3:00 271.7 F 28.7  
4:00 272.4 F 28.6  
  
Min 271.5 F 28.4  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.6 F 0.0  
5:00 233.8 F 0.0  
10:00 230.3 F 0.0  
15:00 229.5 F 0.0  
20:00 228.5 F 0.0  
25:00 233.2 F 0.0  
30:00 225.3 F 0.0  
  
Drying Complete

FILLING: 0:46  
HEATING: 32:22  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:10:23  
CYCLE COMPLETE

06/05/2023 08:37 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_