
Midmark M11 - v1.0.5
Total Cycles: 500

Name: RITTER2
Sterilizer ID: V2313228
10/28/2021 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 183.3 F 4.1
4:00 200.9 F 9.5
6:00 222.4 F 6.8
8:00 233.8 F 10.9
10:00 235.1 F 9.1
12:00 238.7 F 9.7
14:00 249.9 F 15.7
16:00 262.8 F 23.1
18:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.9
1:00 272.6 F 29.4
2:00 271.7 F 29.1
3:00 272.5 F 29.4
4:00 271.5 F 28.8

Min 271.5 F 28.7
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.5 F 0.0
5:00 226.7 F 0.0
10:00 194.4 F 0.0
15:00 192.6 F 0.0
20:00 184.8 F 0.0
25:00 181.5 F 0.0
30:00 179.4 F 0.0
Drying Complete

FILLING: 2:41
HEATING: 18:05
STERILIZING: 4:00
VENTING: 3:02
DRYING: 30:00
TOTAL CYCLE: 00:57:49
CYCLE COMPLETE

10/29/2021 06:38 AM
ACCEPTED: X
REJECTED:
Operator: U02