
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
1.8uS / 1.1ppm
CYCLE NUMBER 000030
12:20 10/08/2020
RUBBER AND PLASTICS
121 C for 15 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 7:49
STERILIZING 8:33
123.4 C 119kPa 8:33
123.4 C 120kPa 9:03
123.4 C 120kPa 9:33
123.3 C 120kPa 10:03
123.4 C 120kPa 10:33
123.4 C 121kPa 11:03
123.3 C 120kPa 11:33
123.1 C 120kPa 12:03
123.5 C 117kPa 12:33
123.3 C 120kPa 13:04
123.3 C 120kPa 13:34
123.4 C 120kPa 14:04
123.3 C 120kPa 14:34
123.5 C 118kPa 15:04
123.3 C 120kPa 15:34
123.5 C 120kPa 16:04
123.2 C 119kPa 16:34
123.4 C 120kPa 17:04
123.5 C 120kPa 17:34
123.5 C 120kPa 18:04
123.4 C 120kPa 18:34
123.4 C 120kPa 19:04
123.3 C 120kPa 19:34
123.4 C 117kPa 20:04
123.4 C 120kPa 20:34
123.0 C 117kPa 21:04
123.4 C 118kPa 21:34
123.4 C 120kPa 22:05
123.4 C 120kPa 22:35
123.4 C 120kPa 23:05
123.5 C 120kPa 23:33
VENTING 23:33

AIR DRYING 24:19
CYCLE COMPLETE 54:19

Digital Signature #
4B5AE01C4BDB533F
