
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003945
8:50 06/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:36
PRESSURIZING 2:42
STERILIZING 3:48
136.4 C 223kPa 3:48
136.3 C 225kPa 4:18
136.3 C 224kPa 4:48
136.0 C 222kPa 5:18
136.0 C 220kPa 5:48
136.2 C 220kPa 6:18
136.2 C 221kPa 6:48
135.9 C 221kPa 7:18
135.9 C 221kPa 7:18
VENTING 7:18
AIR DRYING 8:36
CYCLE COMPLETE 14:03

Digital Signature #
9ED8CE7F7C7F51CC
