
Midmark M11 - v1.0.0
Total Cycles: 2732

Name: RITTER 1
Sterilizer ID: V2178198
01/04/2023 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.9 F 0.0
2:00 191.8 F 0.0
4:00 207.0 F 2.3
6:00 224.5 F 7.4
8:00 236.1 F 12.5
10:00 234.3 F 9.1
12:00 240.5 F 9.8
14:00 245.9 F 13.4
16:00 244.8 F 12.6
18:00 258.3 F 19.7
20:00 269.8 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.5
2:00 272.3 F 28.9
3:00 271.7 F 28.2
4:00 272.6 F 29.0

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 229.8 F 0.0
10:00 237.9 F 0.0
15:00 228.2 F 0.0
20:00 225.7 F 0.0
25:00 227.1 F 0.0
30:00 227.6 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 20:52
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:22
CYCLE COMPLETE

01/04/2023 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____