
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000404
6:58 03/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:59
PRESSURIZING 3:02
STERILIZING 4:03
136.0 C 220kPa 4:03
136.4 C 224kPa 4:33
136.4 C 223kPa 5:03
136.3 C 223kPa 5:33
136.3 C 221kPa 6:03
136.2 C 220kPa 6:33
136.4 C 222kPa 7:03
136.3 C 223kPa 7:33
136.3 C 223kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 8:50

Digital Signature #
C427D425D561B721
