
Midmark M11 - v1.0.5
Total Cycles: 353

Name: RITTER2
Sterilizer ID: V2313228
09/13/2021 08:05 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 163.9 F 1.9
4:00 171.1 F 5.5
6:00 192.3 F 10.1
8:00 209.1 F 6.2
10:00 225.2 F 10.1
12:00 229.0 F 8.0
14:00 241.3 F 13.1
16:00 241.8 F 12.0
18:00 254.5 F 18.5
20:00 265.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.6 F 29.7
2:00 272.2 F 30.2
3:00 272.1 F 29.4
4:00 272.7 F 30.2

Min 271.6 F 29.3
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 205.1 F 0.0
10:00 210.3 F 0.0
15:00 210.5 F 0.0
20:00 190.4 F 0.0
25:00 191.9 F 0.0
30:00 200.4 F 0.0
Drying Complete

FILLING: 1:32
HEATING: 21:44
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:21
CYCLE COMPLETE

09/13/2021 09:07 AM
ACCEPTED: X
REJECTED:
Operator: U01