

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.5uS / 0.3ppm  
CYCLE NUMBER 001505  
15:34 05/10/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:18  
PRESSURIZING 2:27  
STERILIZING 3:24  
136.0 C 220kPa 3:24  
136.4 C 225kPa 3:54  
136.2 C 224kPa 4:24  
136.2 C 221kPa 4:54  
136.2 C 223kPa 5:24  
136.5 C 224kPa 5:54  
136.4 C 225kPa 6:24  
136.4 C 226kPa 6:54  
136.4 C 226kPa 6:54  
VENTING 6:54  
AIR DRYING 7:46  
CYCLE COMPLETE 8:08  
-----

Digital Signature #  
40CB3F179CC43AAD  
-----