
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000205
14:28 21/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:44
STERILIZING 5:15
136.0 C 229kPa 5:15
136.4 C 233kPa 5:45
136.3 C 231kPa 6:15
136.4 C 233kPa 6:45
136.3 C 233kPa 7:15
136.4 C 233kPa 7:45
136.4 C 233kPa 8:15
136.3 C 232kPa 8:45
136.3 C 232kPa 8:45
VENTING 8:45
AIR DRYING 9:47
CYCLE COMPLETE 11:02

Digital Signature #
8048393EC8A51A7F
