
Midmark M11 - v1.0.5
Total Cycles: 2049

Name: RITTER2
Sterilizer ID: V2313228
02/14/2023 02:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.6 F 0.0
2:00 186.2 F 3.3
4:00 204.5 F 8.0
6:00 224.0 F 13.5
8:00 227.8 F 8.9
10:00 236.7 F 10.2
12:00 243.3 F 12.9
14:00 243.3 F 12.3
16:00 256.4 F 19.3
18:00 268.1 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.6 F 29.6
2:00 271.6 F 29.0
3:00 272.6 F 29.7
4:00 272.0 F 29.1

Min 271.5 F 28.9
Max 272.8 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.7 F 0.0
5:00 215.5 F 0.0
10:00 224.2 F 0.0
15:00 224.9 F 0.0
20:00 222.5 F 0.0
25:00 221.6 F 0.0
30:00 220.6 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:15
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:16
CYCLE COMPLETE

02/14/2023 03:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____