
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007892
9:50 26/09/2024

WRAPPED

135 C for 10 min.

CYCLE START	0:00
CONDITIONING	2:06
PRESSURIZING	4:16
STERILIZING	5:04
136.0 C 220kPa	5:04
136.4 C 225kPa	5:35
136.1 C 224kPa	6:05
136.1 C 221kPa	6:35
136.2 C 222kPa	7:05
136.2 C 223kPa	7:35
136.2 C 224kPa	8:05
136.4 C 221kPa	8:35
136.4 C 222kPa	9:05
136.4 C 221kPa	9:35
136.4 C 222kPa	10:05
136.4 C 223kPa	10:35
136.4 C 222kPa	11:05
136.4 C 224kPa	11:35
136.4 C 224kPa	12:05
136.4 C 224kPa	12:35
136.2 C 223kPa	13:05
136.4 C 222kPa	13:35
136.4 C 221kPa	14:05
136.3 C 222kPa	14:35
136.4 C 225kPa	15:05
136.4 C 225kPa	15:05
VENTING	15:05
AIR DRYING	15:57
CYCLE COMPLETE	16:23

Digital Signature #

8DFF241B35AD8508
